



ORIGINAL ARTICLE

Feasibility of optical quality analysis system for the objective assessment of accommodation insufficiency: a phase 1 study

Esther López-Artero^a, Nuria Garzón^{a,b,*}, Manuel Rodríguez-Vallejo^c,
María García-Montero^{a,b}

^a Miranza Group, Madrid, C/Galileo 104, 28003 Madrid, Spain

^b Optics II Department, Faculty of Optics and Optometry, Complutense University of Madrid, Madrid, Spain

^c Qvision, Department of Ophthalmology, Vithas Virgen del Mar Hospital, 04120, Almería, Spain

Received 11 January 2020; accepted 29 June 2020

KEYWORDS

Accommodation
insufficiency;
Amplitude
accommodation;
Double-Pass

Abstract

Purpose: To assess differences in a new objective metric obtained with a double-pass technique between a group with accommodation insufficiency (AI) and a control group and to explore the diagnostic capabilities of this new tool in comparison to conventional procedures.

Methods: Retrospective cross-sectional case-control phase 1 study. Two groups age ranging from 8 to 18 years were recruitment: AI and control group. The diagnostic criterion of AI was based on monocular accommodative amplitude (AA) 2 D below Hofstetter's calculation for minimum AA and monocular accommodative facility (MAF) failing with minus lens and cut-off at ≤ 6 cycles per minute. Accommodative response with a double pass device (HD Analyzer, Visiometrics) was measured, performing an evaluation from +1.00 D to -3.50D (-0.5D steps), offering the width of the profile at 50% (WP) in minutes of arc.

Results: Differences were found between groups for the AA, MAF and MEM retinoscopy ($p < 0.0001$, $p < 0.001$, $p = 0.037$). The discriminative capacity of MEM retinoscopy for AI diagnosis was significant and the cut-off that maximized the sensitivity and specificity was > 0.5 D. Considering WP 50% in different points, the discriminative AI diagnosis capacities for the points of 2.0 D and 2.50 D were significant (ROC-AUC 0.78; $p = 0.03$ and $p = 0.02$).

Conclusions: Double-pass system metric differed between patients with AI and control group, therefore the aim of a Phase I study was achieved. Further steps with higher sample sizes are required to evidence if the system really provides any advantage versus conventional methods in the diagnosis of AI.

© 2020 Spanish General Council of Optometry. Published by Elsevier España, S.L.U. This is an open access article under the CC BY-NC-ND license (<http://creativecommons.org/licenses/by-nc-nd/4.0/>).

* Corresponding author at: Optometry and Vision Department, Faculty of Optics and Optometry, Complutense University of Madrid. Avda Arcos de Jalón, 118, 28037 Madrid, Spain.

E-mail address: ngarzon@ioamadrid.com (N. Garzón).

<https://doi.org/10.1016/j.optom.2020.06.004>

1888-4296/© 2020 Spanish General Council of Optometry. Published by Elsevier España, S.L.U. This is an open access article under the CC BY-NC-ND license (<http://creativecommons.org/licenses/by-nc-nd/4.0/>).

Introduction

Accommodation refers to the change in the dioptric power of the eye. Thus, the light comes focused on the retina from different distances from far to near due to changes in the shape of the crystalline lens.¹ Measurement of accommodative function is mandatory for the management of presbyopia, the physiological loss of amplitude of accommodation with age² in addition to the evaluation of non-presbyopic population with some accommodative dysfunctions such as Accommodative Insufficiency (AI).³

AI is a type of dysfunction in which the amplitude of accommodation (AA) is less than the age-expected value in non-presbyopic eyes.⁴ AI leads to several symptoms such as blurred vision, diplopia, visual fatigue or asthenopia, headache, difficulty to change the focus from distance to near vision, loss of concentration and understanding associated with near activities, such as reading.⁵⁻⁷ Therefore, AI can interfere with academic and educational learning, which is especially important in children and school-age students.⁸

There is a great variability regarding the prevalence of AI in children. As reported in various studies, prevalence rates can range from 2.0%⁹ to 61.7%,¹⁰ depending on the studied population and diagnostic criteria.¹¹ In fact, the prevalence of cases of AI within the same clinical population varies with the clinical diagnostic criteria selected.^{6,12-14} Although there is not a single criterion for the detection of AI, according to the literature, the push-up monocular accommodative amplitude at least 2D below Hofstetter's formula $[15-(0.25 \times \text{age})]$ and failing with the minus flip lens (-2.00 D) in monocular accommodative facility (MAF) seem to be the signs most used for diagnosing AI.¹³ Clinically, AA has usually been measured for more than a century with the conventional push-up method,¹⁵ based on increasing the accommodative demand, moving the test chart towards the patient's eyes until it begins to blur. However, several studies have shown that subjective techniques generally overestimate the measured AA in comparison with objective techniques due to the depth-of-field of the eye, and many factors that can add errors when taking measurement (ie. pupil size,...).¹⁶⁻¹⁸

Although most of the methods used for the measurement of AA in clinical practice are subjective (push-up, push-away, minus lens) or partly objective requiring the clinician's interpretation (monocular estimate method (MEM) retinoscopy), some objective techniques have been reported in the literature to measure the accommodative response.¹⁹⁻²⁶ One of this is the double-pass technique, based on recording images of a point-source object after reflection on the retina and a double pass through the ocular media,²⁵ which has been widely used in daily clinical practice to assess retinal image quality in patients with different eye conditions, and lately has been also applied to the study of accommodation with two investigations which suggest the suitability of the double-pass system to measure accommodation response on healthy adults.^{25,26}

The main aim of this Phase I study was to assess differences in a new objective metric obtained with a double-pass technique between a group with accommodation insufficiency (AI) and a control group. A secondary aim was to explore the diagnostic capabilities of this new tool in comparison to conventional procedures

Material and methods

Subjects

This prospective case-control phase 1 study was conducted at XXX. They were subjects who came to the clinic for the exploration of any kind of functional problem including from blurred vision up to the presence of symptoms associated to daily tasks for which an extended accommodative and binocular evaluation was required. All patients provided written informed consent. The study was approved by the local ethics committee, and it was performed in accordance with the Declaration of Helsinki.

Inclusion criteria were patients with corrected distance visual acuity (CDVA) of 20/20 or better, spherical refractive error ranging from +5 D to -8 D, cylinder refractive error <1.50 D and age ranging from 8 to 18 years. Exclusion criteria included history of ocular disease (eg, uveitis, amblyopia, glaucoma, strabismus), active ocular allergy, refractive surgery, systemic medication known to affect tear film production, any acute or chronic condition or the presence of pupil abnormalities, any binocular or accommodative dysfunction beyond AI and abnormal fundus. Subjects were divided into two groups: AI and control group. For AI diagnosis the following criteria was followed: monocular accommodative amplitude (AA) 2 D diopters below Hofstetter's calculation for minimum amplitude $[15-(0.25 \times \text{age})]$ ²⁷ and monocular accommodative facility (MAF) failing with minus lens and cut-off at ≤ 6 cycles per minute (cpm).^{3,28,29}

Visual examination sequence

All subjects had a previous comprehensive ophthalmic examination including: clinical history, uncorrected distance visual acuity (UDVA), corrected distance visual acuity (CDVA), corrected distance near visual acuity (CDNVA), subjective and cycloplegic refraction, slit-lamp examination of the anterior segment and funduscopy.

Mesopic pupil diameter was measured by OPD-Scan III (Nidek Co., Ltd.). A complete binocular vision and accommodative evaluation was performed with full subjective refraction by the same optometrist (EL). Monocular AA was tested using the push-up method performing with two lines lower than DCNVA³⁰ with ETDRS near visual acuity chart 2000 (Precision Vision, Inc., Woodstock, IL). Monocular accommodative facility (MAF) was conducted following the procedure of Zellers et al.³¹ at 40 cm using ± 2.00 D flip lenses. Accommodative lag was measured by monocular estimate method (MEM) retinoscopy. MEM retinoscopy was performed under moderate room illumination with the patient seated 33 cm from the retinoscope with the Welch Allyn fixation card attached to the head of the retinoscope.^{32,33}

Optical quality analysis system – high definition analyzer

The HD Analyzer™ (Visiometrics S.L., Terrasa, Spain) is based on the double-pass method and provides an objective clinical evaluation of the eye's optical quality through




Figure 1 The right side of the image shows the retinal image intensity distribution profile that provides the HD Analyzer™ software. The profile width values are displayed at 50% and 10% of their maximum. The value of the width of the profile at 50% (WP) in minutes of arc (3.02 min of arc in the image) is the metric of optical quality used in the current pilot and phase I study. The higher value, the worse quality of vision.

150 the Point Spread Function (PSF) measurement. The HD
151 Analyzer™ software shows the retinal image intensity dis-
152 tribution profile and calculates and registers the value of
153 the width of the profile at 50% (WP) in minutes of arc. With
154 these values the clinician can obtain information about the
155 size of the image profile. The higher value, the worse quality
156 of vision (Fig. 1).

157 Measurement of accommodative response

158 Patients were tested with the double-pass system after com-
159 prehensive ophthalmic examination and complete binocular
160 vision and accommodative evaluation under non cycloplegic
161 condition.

162 The HD Analyzer™ has a software function named
163 "Pseudoaccommodation". Previously starting this software
164 function, the subject's spherical refractive error was auto-
165 matically corrected by the Badal system and patient's
166 astigmatism >0.50 D was corrected by means of trial lenses
167 placed in the front part. Measurement process of accom-
168 modative range is based on a sweep of four diopters [from
169 +1D to -3D] around the selected value, usually of the best
170 spherical correction. The device makes an evaluation in
171 -0.5D steps taking double-pass images in each position until
172 reach the final position and register retinal image intensity
173 distribution profiles and offer the WP (optical quality) value
174 at each accommodative vergence, expressed in minutes of
175 arc (Fig. 2). For each different accommodation states an
176 image of the stimulus was acquired when the change of
177 vergence was compensated by the Badal system.

178 Measurements were taken under mesopic conditions
179 (3-3.5 lux) measured by a photometer (Gossen Starlite
180 II Foto- und Lichtmesstechnik GmbH, Nürnberg) and the
181 recording of accommodative response started after a 5-
182 minute period of adaptation to the dark. Subjects were
183 seated at the device with their head stabilized in the
184 instrument's chin rest and forehead strap. Before start the
185 recording, patients were asked to blink few times and to
186 keep the target focused. Once the measurement of accom-
187 modation started the latter request was reinforced through
188 all the process. The patient was instructed to try to focus
189 the visual target (landscape with a house) that appeared
190 inside the instrument. The measurement was performed

191 with a 5 mm artificial pupil included in the device in non-
192 cycloplegic conditions.

193 Statistical analysis

194 Despite both eyes were measured, only right eyes were
195 included in the statistical analysis. The normality of data
196 distributions for the variables included in the study was
197 tested with the Saphiro-Wilk test. Descriptive statistics are
198 detailed in the results section as mean ± standard deviation.
199 Differences between groups (AI vs control) were tested with
200 the *t*-test for independent samples and the Receiver Operat-
201 ing Characteristic (ROC) curves were computed considering
202 AI diagnosis as the state variable (AA < 2 D normative by age
203 and MAF failing with minus lens and cut-off at ≤ 6 cpm). The
204 Youden index was calculated for determining the diagnostic
205 cut-off values for all the variables involved in the diagno-
206 sis of the AI.³⁴ The area under the curves (AUC) for the WP
207 (From 0 to -2.50 D) was computed with the *trapz* function of
208 MATLAB (R2018Ba; MathWorks, Natick, MA) after subtracting
209 at each point the WP at the 0 D location (WP_x-WP₀). SPSS
210 software (version 24; SPSS Inc., Chicago, IL) was used for
211 the statistical analysis.

212 Results

213 Eighteen subjects (7 men and 11 women) were recruited.
214 Eight of them met the inclusion criteria of IA group and 10
215 were included in the control group. Table 1 shows that no
216 differences in age or mesopic pupil diameter were found
217 between groups (*p* > 0.05). The mesopic pupil size was rang-
218 ing from 5.65 to 8.43 mm in such a way the pupil was higher
219 than the artificial 5 mm pupil used during the WP mea-
220 surement in all the cases. Differences were found between
221 groups for the AA, MAF and MEM retinoscopy (*p* < 0.05)
222 (Table 1).

223 Regarding HD Analyzer, despite the device measured the
224 range from +1.00 D to -3.50 D only the range from 0 to
225 -2.50 D was finally included in the analysis because the
226 device could not obtain a numeric value in 4 eyes in the
227 -3.00 D defocus and in 11 eyes in the -3.50 D defocus. The
228 Fig. 2 shows the WP for each one of the defocus lens tested,




Figure 2 Example of the 0.00 D, 0.5 D, 1.0 D, 1.5 D, 2.0 D and 2.5 D double-pass images recorded during the sequence of measurement of accommodative response in an AI subject. For each point tested a double image is represented. Although the entire PSF is not observed in the 2.5 D image, that is only a representation. If that error would occur, many of the measurements obtained would be similar if a limit would have been established. In the upper left part of the upper image, a number is shown that corresponds to the value, in minutes of arc, of the width of the profile at 50% (WP) in minutes of arc (2.30 min of arc for 0.0 D, 5.54 min of arc for 0.5 D, ...31.34 min of arc for 2.5 D). In the bottom image, a simulation of how the image projected onto the retina of an optotype (Snellen letter E) is shown.

Table 1 Descriptive statistics and comparative analysis results from AI group and control group.

	AI group (n = 8) Mean value ± SD	Control group (n = 10) Mean value ± SD	t -test
Age	12.88 ± 3.56	14.40 ± 3.06	0.977, p = 0.34
Mesopic pupil diameter (mm)	7.08 ± 0.90	6.90 ± 0.66	-0.49, p = 0.63
Push-up AA (D)	4.90 ± 2.89	12.53 ± 2.05	6.56, p < 0.0001**
Theoretical AA - Push-up AA (D)	6.88 ± 2.81	-1.13 ± 1.75	-7.42, p < 0.0001**
MAF (cpm)	4.13 ± 2.75	14.00 ± 2.54	7.91, p < 0.0001**
MEM retinoscopy (D)	0.88 ± 0.76	0.05 ± 0.76	-2.28, p = 0.037*
HD Analyzer			
AUC for the WP	41.45 ± 19.67	24.40 ± 9.44	-2.52, p = 0.049*
WP (minutes of arc) at defocus of:			
0.5 D	1.39 ± 1.73	1.18 ± 1.28	-0.30, p = 0.77
1.0 D	5.53 ± 4.38	3.34 ± 2.54	-1.33, p = 0.20
1.5 D	9.96 ± 6.51	5.08 ± 2.78	-2.15, p = 0.047*
2.0 D	15.92 ± 6.79	9.46 ± 4.96	-2.34, p = 0.033*
2.5 D	17.31 ± 8.16	10.70 ± 7.43	-1.80, p = 0.09

AI: Accommodative Insufficiency; AA: accommodative amplitude; Theoretical AA: Hofstetter's minimum value [15-(0.25×age)]; MAF: monocular accommodative facility; cpm: cycles per minute; MEM: monocular estimate method; AUC for the WP: area under the through focus width profile measurement of HD Analyzer; WP: width of the profile at 50%.

* p < 0.05.
** p < 0.005.

obtaining significant differences for the points of -1.50 D and -2.00 D. The AUC for the WP was also significantly different between groups (Table 1).

The diagnostic performance for MEM retinoscopy, AUC for the WP and WP values for each one of the defocus lens tested was determined and data were summarized for area under the ROC curve (ROC-AUC), standard error of the ROC-AUC, cut-off value, sensitivity, specificity and Youden index in Table 2. The discriminative capacity of MEM retinoscopy for AI diagnosis was significant (ROC-AUC 0.77; P < 0.05) and the cut-off that maximized the sensitivity and specificity was > 0.5 D. For the consideration of a singular defocus point in the measurement of the WP, instead of the integration of all with the AUC for the WP, the points with highest Youden indices were the 2.0 D and 2.5 D defocus locations. The dis-

criminative AI diagnosis capacities of WP for the points of 2.0 D and 2.50 D were significant (ROC-AUC 0.78; P < 0.05) (Table 2) (Fig. 3).

Discussion

Subjective methods to measure accommodation functions are based in the ability of the accommodative system to detect blur, but it is mandatory the active participation of the subjects in order to keep attention in the target to detect and inform about blur. Objective methods have the advantage to not depend so stronger on the subject's response, even more in the children population. Diagnostic studies are mandatory in order to evidence the most useful procedures, that alone or in combination with others, can

Table 2 Receiver operating characteristic (ROC) analysis for the diagnosis of accommodation insufficiency.

	ROC-AUC	SE	p-value	Cut-off	Sensitivity	Specificity	Youden I
MEM retinoscopy (D)	0.77	0.13	0.04*	> 0.5	87.5	80	0.67
HD Analyzer							
AUC for the WP	0.76	0.12	0.03*	>37.49	50	100	0.5
WP (minutes of arc) at defocus of:							
0.5 D	0.58	0.15	0.63	≥0.41	50	80	0.3
1.0 D	0.65	0.14	0.28	>1.32	100	30	0.3
1.5 D	0.74	0.14	0.08	>5.9	87.5	60	0.48
2.0 D	0.78	0.12	0.03*	>11.5	75	80	0.55
2.5 D	0.78	0.12	0.02*	>9.42	100	60	0.6

MEM: monocular estimate method; AUC for the WP: area under the with profile measurement of HD Analyzer; WP: width of the profile at 50% ; ROC-AUC: Area under the ROC curve; SE: standard error of the ROC-AUC; Cut-off: value that maximize the sensitivity and specificity computed through the Youden index.

* p < 0.05.




Figure 3 Representation the WP for each point of the defocus lens tested (0.0 D, 0.5 D, 1.0 D, 1.5 D, 2.0 D and 2.5 D) for the AI group (dashed line) and control group (continuous line). Ordinate axis represented the difference of WP at the 0 D location (WP₀) between each point (WP_x). The figure shows significant differences between groups (AI vs control) for the points of -1.50 D and -2.00 D. *p < 0.05.

help to the clinician in the detection of anomalies. The Phase I of diagnostic studies involves the inclusion of equal size low sample groups in a cross-sectional case-control design to evaluate whether the test discriminates between patients who clearly have the anomaly.³⁵ Therefore, the main aim of this Phase I study was to evaluate if the HD Analyzer could differentiate subjects with AI from a control group.

In the current study, we proposed for AI diagnosis criterion the subjective push-up monocular AA 2 D below Hofstetter's calculation for minimum amplitude and MAF failing with minus lens with cut-off at ≤ 6 cpm. Despite nowadays there is no common clinical diagnostic criterion for the detection of AI all of them incorporate the push-up monocular accommodative amplitude at least 2 D below Hofstetter's calculation for minimum amplitude.^{3,4,12,28,29,36-39}

Regarding measuring AA, various studies showed that the subjective push-up monocular AA method overestimated the objective AA,^{16,19,22} particularly in young children.¹⁹ There-

fore, there are several distinct sources of error that affect the measure: eye's depth-of-field, target size, illumination, end-point criteria, proximal cues, subject variability affect the outcome of the AA measurement and pupil size.¹⁶⁻¹⁸ Our sample did not show differences in mesopic pupil diameter between groups and also the mesopic pupil size was ranging from 5.65 to 8.43 mm in such a way the pupil was higher than the artificial 5 mm pupil used during the WP measurement in all the cases.

Nowadays, the subjective push-up test is the method most routinely used to measure accommodative amplitude but it is inadequate for demonstrating true accommodative amplitude. Wold et al. suggested that objective measures of accommodation should be used to determine whether accommodation can be restored in presbyopia in case of scleral-expansion surgical procedures or replacement of the cataractous crystalline lens with so-called accommodating intraocular lenses.¹⁶ On the other hand, in the case of non-presbyopic subjects, specific children, our target population, Anderson et al. reported relatively stable amplitude does not begin to decline rapidly until the third decade of life.⁴⁰ Therefore, the AA-age relation was not a bias in the current study, even observing that there weren't differences in age between groups.

It is well known that objective techniques are more accurate to determine the accommodative ability^{16,19} and several methods have been reported in the literature based on autorefractometry, aberrometers and double pass technologies.^{19-26,41} In fact, global image quality-based measurements seem more appropriate than techniques based just on defocus measurements in order to control spherical aberration (major contributor to higher-order aberrations with the accommodation).⁴² However, as far as we know, there are not researches that use objective methods to study the accommodative response in non-presbyopic subjects with AI. So, to the best of our knowledge, this is the first study to analyzed the accommodative response using an objective technique based on an optical quality metric using double-pass system: the width of the retinal image intensity distribution profile at 50% (WP) in minutes of arc (Fig. 1).

The double-pass technique, based on recording images of a point-source object after reflection on the retina and a double pass through the ocular media²⁵ has been widely

used in daily clinical practice to assess retinal image quality in patients with different eye conditions, and lately has been also applied to the study of accommodation with two investigations which suggest the suitability of the double-pass system to measure accommodation response on healthy adults.^{25,26}

The metric of optical quality used in the current pilot and phase I study was based on the retinal image intensity distribution profile and the value of the width of the profile at 50% (WP) in minutes of arc was the study variable. Instead, other authors used the volume under the Modulation Transfer Function.²⁶

The WP was analyzed for 0 D, 0.5 D, 1.0 D, 1.5 D, 2.0 D and 2.5 D and the ROC-AUC calculated for these defocus locations was higher for the AI group than for the control group. However, the differences were not significant between groups for defocus locations ≤ 1.5 D. (Fig. 2) (Table 1) The diagnostic performance based on ROC analysis for these significant variables (AUC for the WP and WP values from 1.50 D defocus lens tested) was significant for AUC for the WP and for WP value from 2.0 D. (Table 2) In order to establish a diagnostic value based on WP, we suggest choosing the singular defocus point in the measurement of the WP, specifically from 2.00 D, instead of the integration of all defocus point with the AUC for the WP. Therefore, the current phase one study would suggested the WP for 2.0 D defocus point as a possible clinical sign of AI, setting the cutoff point at 11.5 min of arc. (Table 2)

Moreover other authors have reported changes from other distances, such as the case of Chase et al.²⁰ and Franco et al.,²³ although they did not used double-pass system. Chase et al.²⁰ used an open-field autorefractor, Grand Seiko autorefractor (WAM-5500), to measure the stimulus-response functions objectively in college students (mean age 20 years old), with and without near work induced discomfort symptoms as the AI criterion diagnosis. They reported significant correlations at viewing distances of 4D or closer. Instead, the device and the AI diagnostic criterion were different that ours. On the other hand, Franco et al.²³ reported change in accommodation functions measured by Hartmann-Shack aberrometer, based on retinal image. They concluded that measurement of wavefront ocular aberrations can be a tool to diagnose accommodative disorders; however they reported only one case of a presbyopic subject while we studied young population.

Another objective technique to measure the accommodative response is MEM retinoscopy and some authors include it for AI criterion.^{29,40} The MEM retinoscopy was different between groups in the current study and the diagnostic performance based on ROC analysis was significant with a cut off higher than 0.50 D. Therefore, our result suggests that MEM retinoscopy could be a clinical sign to consider in the diagnosis of AI, setting the cut-off point at 0.50 D. Contrary, these results are not in accordance with those presented by Cacho et al.³ and they recommended use the standard deviation as the cutoff to maintain a high level of specificity.

It is relevance to reinforce that this technique is considered objective because it does not depend on the subject's response, but it depends on the examiner. Rouse et al. demonstrated its validity and established normative data³² in 0.75 D, however the mean differences between measurement found in a repeatability study were 0.23 ± 0.50 D.⁴³

Consequently, despite the statistically significant results, it is not clinically relevant to consider a MEM value higher than 0.50 D as a clinical sign of AI when the measure itself has an error of the same value.

Although the main aim of a Phase I study was achieved detecting significant differences between groups, it is important to note that the device and the study have some limitations that should be taken into account for the development of the required future studies. Among device limitations, difficulties in the measurement can appear in some subjects such as subjects with abnormal fundus,⁴⁴ myopia induction by the device might affect to the IA detection,⁴⁵ the spatial frequency of the target used by the device might not allow participants to accurately perceive the blur and therefore modifying the accommodative response,⁴⁶ and finally the diagnosis with the device intends to detect AI through the decrease in optical quality with the increase of accommodative demand which differs from the conventional method of measuring the accommodation amplitude. Although the system cannot reach in some cases the accommodative limit, a decrease in optical quality might appear inside the range measured by the device that might help to overcome this limitation detecting the AI through the area under the curve instead of in the depth of field measured range in diopters. However, this metric is not familiar for clinicians and needs of cut-off points specifically designed for this metric and purpose. Among study limitations, we explored the sensitivity, specificity and cut-off point for detection of the IA as a secondary aim. However, it is important to note that our study has not enough sample size to transfer these values to the clinical practice. To achieve reliable values for this secondary aim, further steps (Phase II, III and IV studies) are required increasing the sample size and including patients with different degrees of the dysfunction.^{35,47}

Conclusion

In conclusion, the study passed the Phase I suggesting that it is therefore worthwhile developing next stages, that can provide reliable evidence about the sensitivity, specificity and cut-off points for IA detection. However, our pilot study does not provide sufficient evidence to conclude that the system really provides any advantage versus conventional methods in the diagnosis of AI. A Phase II study with a higher sample size is required to assess if patients with certain test results are more likely to have AI than patients with other test results.

Conflicts of interest

The authors have no conflicts of interest to declare.

References

1. Rutstein RP, Daum DK. *Anomalies of binocular vision: Diagnosis and management*. 2nd ed. St. Louis: Mosby; 1998:61-65.
2. Schachar RA. The mechanism of accommodation and presbyopia. *Int Ophthalmol Clin*. 2006;46:39-61.
3. Cacho P, García Á, Lara F, Seguí M. Diagnostic signs of accommodative insufficiency. *Optom Vis Sci*. 2002;79:614-620.

- 435 4. Borsting E, Rouse MW, Deland PN, et al. Association of symptoms and convergence and accommodative insufficiency in school-age children. *Optometry*. 2003;74:25-34. 501
- 436 5. Abdi S, Brautaset R, Rydberg A, Pansell T. The influence of 502
- 437 accommodative insufficiency on reading. *Clin Exp Optom*. 503
- 438 2007;90:36-43. 504
- 439 6. Daum KM. Accommodative insufficiency. *Am J Optom Physiol*. 505
- 440 1983;60:352-359. 506
- 441 7. Garcia-Munoz A, Carbonell-Bonete S, Cacho-Martinez P. Symptomatology associated with accommodative and binocular 507
- 442 vision anomalies. *J Optom*. 2014;7:178-192. 508
- 443 8. Hashemi H, Khabazkhoob M, Nabovati P, et al. Accommodative 509
- 444 insufficiency in a student population in Iran. *J Optom*. 510
- 445 2019;12:161-167. 511
- 446 9. Scheiman M, Gallaway M, Coulter R, et al. Prevalence of vision 512
- 447 and ocular disease conditions in a clinical pediatric population. 513
- 448 *J Am Optom Assoc*. 1996;67:193-202. 514
- 449 10. Abdi S, Rydberg A. Asthenopia in schoolchildren, orthoptic 515
- 450 and ophthalmological findings and treatment. *Doc Ophthalmol*. 516
- 451 2005;111:65-72. 517
- 452 11. Cacho P, Garcia A. Do we really know the prevalence of accommodative and nonstrabismic binocular dysfunctions? *J Optom*. 518
- 453 2010;3:185-197. 519
- 454 12. Garcia-Montero M, Antona B, Barrio AR, Nieto-Zayas C, 520
- 455 Martinez-Alberquilla I, Hernandez-Verdejo JL. The role of clinical 521
- 456 diagnosis criteria on the frequency of accommodative 522
- 457 insufficiency. *Int J Ophthalmol*. 2019;12:647-653. 523
- 458 13. Cacho P, Garcia A. Is there any evidence for the validity of 524
- 459 diagnostic criteria used for accommodative and nonstrabismic 525
- 460 binocular dysfunctions? *J Optom*. 2014;7:2-21. 526
- 461 14. Hokoda SC. General binocular dysfunctions in an urban optometry 527
- 462 clinic. *J Am Optom Assoc*. 1985;56:560-562. 528
- 463 15. Duane A. An attempt to determine the normal range of accommodation at various ages, being a revision of Donder's experiments. *Trans Am Ophthalmol Soc*. 1908;11:634-641. 529
- 464 16. Wold JE, Hu A, Chen S, Glasser A. Subjective and objective 530
- 465 measurement of human accommodative amplitude. *J Cataract Refract Surg*. 2003;29:1878-1888. 531
- 466 17. Atchison DA, Capper EJ, McCabe KL. Critical subjective measurement of amplitude of accommodation. *Optom Vis Sci*. 1994;71:699-706. 532
- 467 18. Burns D, Allen P, Edgar D, Evans B. Sources of error in clinical 533
- 468 measurement of the amplitude of accommodation. *J Optom*. 534
- 469 2019. 535
- 470 19. Anderson HA, Stuebing KK. Subjective versus objective accommodative amplitude: Preschool to presbyopia. *Optom Vis Sci*. 2014;91:1290-1301. 536
- 471 20. Chase C, Tosha C, Borsting E, Ridder WH 3rd. Visual discomfort and objective measures of static accommodation. *Optom Vis Sci*. 2009;86:883-889. 537
- 472 21. Tarrant J, Roorda A, Wildsoet CF. Determining the accommodative response from wavefront aberrations. *J Vis*. 2011;10:4. 538
- 473 22. López-Gil N, Fernández-Sánchez V, Thibos LN, Montés-Micó R. Objective amplitude of accommodation computed from optical quality metrics applied to wavefront outcomes. *J Optom*. 2009;2:223-234. 539
- 474 23. Franco S, Gomes J. Real-time measurement of ocular wavefront aberrations in symptomatic subjects. *Biomed Res Int*. 2018;2018. 540
- 475 24. Zhou X-T, Wl X-Y, Zhou X-T, Yu Z-Q. Wavefront aberration changes caused by a gradient of increasing accommodation stimuli. *Eye*. 2015;29:115-121. 541
- 476 25. Aldaba M, Vilaseca M, Diaz-Douton F, Arjona M, Pujol J. Measuring the accommodative response with a double-pass system: Comparison with the Hartmann-Shack technique. *Vision Res*. 2012;62:26-34. 542
- 477 26. Aldaba M, Vilaseca M, Arjona M, Pujol J. Age-related changes in accommodation measured with a double-pass system. *Ophthalmic Physiol Opt*. 2013;33:508-515. 543
- 478 27. Hofstetter HW. Useful age-amplitude formula. *Optom World*. 1950;38:42-45. 544
- 479 28. Lara F, Cacho P, Garcia A, Megias R. General binocular disorders: Prevalence in a clinic population. *Ophthalmic Physiol Opt*. 2001;21:70-74. 545
- 480 29. Porcar E, Martinez-Palomera A. Prevalence of general binocular dysfunctions in a population of university students. *Optom Vis Sci*. 1997;74:111-113. 546
- 481 30. Carlson NB, Kurtz D. *Clinical procedures for ocular examination*. 4th ed. New York: Mac Graw-Hill; 2016. 547
- 482 31. Zellers JA, Alpert TL, Rouse MW. A review of the literature and a normative study of accommodative facility. *J Am Optom Assoc*. 1984;55:31-37. 548
- 483 32. Rouse MW, Hutter RF, Shiftlett R. A normative study of the accommodative lag in elementary school children. *Am J Optom Physiol Optics*. 1984;61:693-697. 549
- 484 33. Locke LC, Somers W. A comparison study of dynamic retinoscopy techniques. *Optom Vis Sci*. 1989;66:540-544. 550
- 485 34. Youden WJ. Index for rating diagnostic tests. *Cancer*. 1950;3:32-35. 551
- 486 35. Knottnerus JA, Muris JW. Assessment of the accuracy of diagnostic tests: The cross-sectional study. *J Clin Epidemiol*. 2003;56:1118-1128. 552
- 487 36. Sterner B, Abrahamsson M, Sjostrom A. Accommodative facility training with a long term follow up in a sample of school aged children showing accommodative dysfunction. *Doc Ophthalmol*. 1999;99:93-101. 553
- 488 37. Dwyer. The prevalence of vergence accommodation disorders in a school-age population. *Clin Exp Optom*. 1992;75:10-18. 554
- 489 38. Marran LF, De Land PN, Nguyen AL. Accommodative insufficiency is the primary source of symptoms in children diagnosed with convergence insufficiency. *Optom Vis Sci*. 2006;83:281-289. 555
- 490 39. Rouse MW, Borsting E, Hyman L, et al. Frequency of convergence insufficiency among fifth and sixth graders. *Optom Vis Sci*. 1999;76:643-649. 556
- 491 40. Anderson HA, Hentz G, Glasser A, Stuebing KK, Manny RE. Minus-lens-stimulated accommodative amplitude decreases sigmoidally with age: A study of objectively measured accommodative amplitudes from age 3. *Invest Ophthalmol Vis Sci*. 2008;49:2919-2926. 557
- 492 41. Win-Hall DM, Ostrin LA, Kasthurirangan S, Glasser A. Objective accommodation measurement with the Grand Seiko and Hartinger coincidence refractometer. *Optom Vis Sci*. 2007;84:879-887. 558
- 493 42. Buehren T, Collins MJ. Accommodation stimulus-response function and retinal image quality. *Vision Res*. 2006;46:1633-1645. 559
- 494 43. Antona B, Sanchez I, Barrio A, Barra F, Gonzalez E. Intra-examiner repeatability and agreement in accommodative response measurements. *Ophthalmic Physiol Opt*. 2009;29:606-614. 560
- 495 44. Lee K, Sohn J, Choi JG, Chung SK. Optical quality in central serous chorioretinopathy. *Invest Ophthalmol Vis Sci*. 2014;55:8598-8603. 561
- 496 45. Rosenfield M, Ciuffreda KJ, Ong E, Azimi A. Proximally induced accommodation and accommodative adaptation. *Invest Ophthalmol Vis Sci*. 1990;31:1162-1167. 562
- 497 46. Charman WN, Tucker J. Dependence of accommodation response on the spatial frequency spectrum of the observed object. *Vision Res*. 1977;17:129-139. 563
- 498 47. Bujang MA, Adnan TH. Requirements for minimum sample size for sensitivity and specificity analysis. *J Clin Diagn Res*. 2016;10:ye01-ye6. 564
- 499 565
- 500 566